

ILS Laboratories
8222 Vickers St, Suite 106, San Diego, CA 92111
(619) 329-3999 | ils-lab.com

BPC-157 - 10mg

Tested for: Elyria Bio

PASS



Scan to verify
authenticity at ils-lab.com
Access Code: 4FT5KXQ2

COA #: COA-2026-955273
Lot Number: 030284
Labeled Content: 10mg
Sample Matrix: Lyophilized

Analysis Date: 08/07/2026
Appearance: Good
Volume: 3mL
Date Received: 07/16/2026

Method: Full QC Panel

Identity

BPC-157


Peptide Purity

99.72%

Full QC Panel

Analyte	Specification	Result	Unit	Status
Peptide Purity (HPLC)	>= 95.0%	99.72%	%	PASS
Net Peptide Content	Report Only	10.7	mg	N/A
Identity (HPLC-RTM)	BPC-157	Confirmed	-	PASS

HPLC Chromatogram



BPC-157 10mg - 030284: UV Chromatogram

Heavy Metals Analysis (ICP-MS)

Test	Specification	Result	Status
Arsenic (As)	NMT 1.5 ppm	0.0378	PASS
Cadmium (Cd)	NMT 0.5 ppm	0.0289	PASS
Lead (Pb)	NMT 1.0 ppm	0.0279	PASS
Mercury (Hg)	NMT 1.5 ppm	0.929	PASS
Chromium (Cr)	NMT 10.0 ppm	0.0288	PASS

Sterility Testing (PCR)

Test	Specification	Result	Status
Sterility (PCR)	No Growth	No Growth	PASS



Dr. Greg Kalyuzhny

Dr. Greg Kalyuzhny
Lab Director
8/7/2026

COA #: COA-2026-955273
Access Code: 4FT5KXQ2
Verify: portal.ils-lab.com/verify/EccT-Ei875_DeEJH
Issued: 8/7/2026

Endotoxin Testing (USP <85>)

Test	Specification	Result	Status
Endotoxin (USP <85>)	Report Result	0.101 EU/mL	Reported
About this result: Endotoxin is reported as a quantitative value. Acceptable limits vary by product type and matrix, so no universal pass/fail threshold applies to RUO products. This result is below commonly referenced endotoxin thresholds.			

Notes & Methodology

1. Date Tested: 08/07/2026. Methods: Full QC Panel.
2. The sample was confirmed to be BPC-157 by HPLC. Identification by chromatographic retention time comparison with a reference standard.
3. Elemental impurities analyzed by ICP-MS per USP <233> methodology. Acceptance criteria are internal laboratory quality screening limits for research-use materials and do not represent evaluation against any specific pharmacopeial monograph or product specification.
4. Endotoxin tested per USP <85> kinetic turbidimetric method. Acceptance criteria per client specification.
5. Peptide purity determined by RP-HPLC area normalization at 214 nm. Value represents the percentage of the target peptide relative to all peptide-related peaks. Non-peptide process-related impurities, if detected, are excluded from the calculation.




Dr. Greg Kalyuzhny
Lab Director
8/7/2026

COA #: **COA-2026-955273**
Access Code: **4FT5KXQ2**
Verify: portal.ils-lab.com/verify/EccT-Ei875_DeEJH
Issued: 8/7/2026